

External Clock Details for Siglent Spectrum Analyzers (SSA3X, SSA3X Plus, SVA1X, and SSA3X-R Series)

March 25, 2021

1-5Vp, 10MHz +/- 100 ppm (Sine preferred)

This indicates that the external clock can be “off” from 10 MHz by +/- 100 ppm. All frequency measurements are scaled according to the clock. If the clock is not 10 MHz, then the frequencies indicated on the instrument will also be off.



North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free: 877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel: + 86 755 3661 5186
Fax: + 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens