

Calibration Traceability

October 23, 2017

Excerpt from the ANSI-ASQ National Accreditation Board:

<http://anab.org/news/newsletter/winter-2014/nist-traceability-statements>

NIST Traceability Statements

Many calibration customers request on their accredited calibration certificates a statement that the measurement is traceable to the National Institute of Standards and Technology (NIST). They want assurance that the standards used for the calibration are ultimately traceable to the International System of Units (SI) through NIST. The only problem with that request is that in many cases the calibration laboratory doesn't know if NIST traceability exists.

Are you shocked? Not if you understand traceability in an increasingly more sophisticated world. The first important fact the calibration customer needs to know is that NIST may not be the only source of traceability to the SI! Many countries have a national metrology institute (NMI) such as NIST that provides traceability to the SI for certain parameters. Accredited calibration laboratories are not required to attain traceability through a particular NMI as long as the NMI used is a signatory of the International Committee for Weights and Measures (CIPM) Mutual Recognition Arrangement.

The second important fact the calibration customer needs to know is that the calibration laboratory providing an accredited calibration may not know which NMI provided the first step in the traceability chain – but it doesn't matter! An accredited calibration certificate is proof that the calibration laboratory used standards traceable to the SI for a given parameter. Note we didn't say – and this is important – “standards that are calibrated by the NMI” or “standards that are calibrated by NIST.” Very few standards are actually calibrated by an NMI and even fewer are calibrated by NIST. Most standards are calibrated by one or more intermediate calibration laboratories that are themselves accredited and traceable to an NMI.

Accreditation ensures traceability by requiring that each link in the traceability chain from the SI be accredited or that proof of traceability be otherwise proven. Thus, the traceability flows from the SI through an NMI and then through any intermediate laboratories to the customer's calibration provider, and on to the customer's accredited equipment. The further the traceability chain extends from the NMI, the less likely the possibility that the calibrating laboratory knows which NMI provided the original traceability. But because traceable calibrations were required throughout the chain, traceability to the SI is still assured.

A word about NIST numbers: They're not acceptable as proof of traceability! The accredited certificate of calibration, however, should be all the proof an informed customer needs.

Typical SIGLENT Calibration Certificate:



SAMPLE Certificate of Calibration

Certificate No.:

Manufacturer: SIGLENT TECHNOLOGIES CO., LTD

Product Type: Digital Storage Oscilloscope

Model: XXX

Serial NO.: XXX

Calibrated /Test Item: Appearance Test, Function Test, Performance Test and Safety Test

Temperature: 23±5℃

Humidity: < 70% RH

Calibrated Date: XXX

Traceability Information: SIGLENT certifies that the above product meets or exceeds published measurement specifications and has been calibrated using standards traceable to National Metrology Institutes(NIM, NIST, NPL, PTB) that are linked to the international system of units(SI).The policies and procedures used at Siglent facility are based on ISO9001/IEC 17025:2005. This certificate shall not be reproduced, except in full, without the written approval of the calibration facility.

Calibration Equipment:

Model	Description	Serial NO.	Calibration Due Date
R&S SMB100A	Signal Generator	1406.6000K02	11-FEB-2018
Keysight B2901A	Precision Source/Measure Unit	MY51140865	21-JUL-2017
Agilent 34401A	Digital multimeter	MY50390009	11-FEB-2018
Agilent U2004A	Power Meter	MY55320012	12-FEB-2018

Calibration Technician:

Checked by:

Note: The certificate applies only to this Oscilloscope.
Without permission, may not be part of the copy of the certificate.

Here is a PDF copy: [ScopeSampleCalCert](#)



North American Headquarters

SIGLENT Technologies America, Inc
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free: 877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES EUROPE GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel: +86 755 3661 5186
Fax: +86 755 3359 1582
sales@siglent.com
www.siglent.com/ens